

# FOREIGN COUNTRY LABORATORY REVIEW

REVIEW DATE

11-01-2001

NAME OF FOREIGN LABORATORY

Institute of Hygiene and Veterinary Public Health,  
NSVA of MAF

FOREIGN GOV'T AGENCY  
National Sanitary Veterinary Agency  
(NSVA)

CITY & COUNTRY  
Bucharest, Romania

ADDRESS OF LABORATORY  
Campul Mosilor St., No. 5, Sect. 2, Bucharest 73341

NAME OF REVIEWER  
Dr. S.P. Singh

NAME OF FOREIGN OFFICIAL  
Dr. Sergiu Meica, Director

Residue Code/Name			100	111	200	203	400	500	800	902	SAL	EC		
SAMPLING PROCEDURES	REVIEW ITEMS	ITEM #	EVALUATION CODE											
	Sample Handling	01		A	A	A	A	A	A	A	A	A		
	Sampling Frequency	02		A	A	A	A	A	A	A	A	A		
	Timely Analyses	03		A	A	A	A	A	A	A	A	A		
	Compositing Procedure	04		O	O	O	O	O	O	O	O	O		
	Interpret Comp Data	05		O	O	O	O	O	O	O	O	O		
	Data Reporting	06		A	A	A	A	A	A	A	A	A		
ANALYTICAL PROCEDURES	Acceptable Method	07	EVALUATION CODE	A	A	A	A	A	A	A	A	A		
	Correct Tissue(s)	08		A	A	A	A	A	A	A	A	A		
	Equipment Operation	09		A	A	A	A	A	A	A	O	O		
	Instrument Printouts	10		A	A	A	A	A	A	A	O	O		
QUALITY ASSURANCE PROCEDURES	Minimum Detection Levels	11	EVALUATION CODE	A	A	A	A	A	A	A	O	O		
	Recovery Frequency	12		A	A	A	A	A	A	A	O	O		
	Percent Recovery	13		A	A	A	A	A	A	A	O	O		
	Check Sample Frequency	14		A	A	A	A	A	A	A	A	A		
	All analyst w/Check Samples	15		A	A	A	A	A	A	A	A	A		
	Corrective Actions	16		A	A	A	A	A	A	A	A	A		
	International Check Samples	17		N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A		
REVIEW PROCEDURES	Corrected Prior Deficiencies	18	EVAL. CODE	A	A	A	A	A	A	A	A	A		
OTHER REVIEW		19	EVAL. CODE											
		20	EVAL. CODE											

SIGNATURE OF REVIEWER

DATE

<b>FOREIGN COUNTRY LABORATORY REVIEW</b> <i>(Comment Sheet)</i>		<b>REVIEW DATE</b> 11-01-2001	<b>NAME OF FOREIGN LABORATORY</b> Institute of Hygiene and Veterinary Public Health, NSVA of MAF
<b>FOREIGN GOV'T AGENCY</b> National Sanitary Veterinary Agency (NSVA)	<b>CITY &amp; COUNTRY</b> Bucharest, Romania	<b>ADDRESS OF LABORATORY</b> Campul Mosilor St., No. 5, Sect. 2, Bucharest 73341	
<b>NAME OF REVIEWER</b> Dr. S.P. Singh		<b>NAME OF FOREIGN OFFICIAL</b> Dr. Sergiu Meica, Director	

RESIDUE CODES	ITEM NO.	COMMENTS
		Comments: 400 = Trace Elements; Arsenic (AS) was not being tested in previous audit. Instruments were obtained and tests for Arsenic will start from January 2002.